Nikon

Stereoscopic SMZ-2B Microscope

INSTRUCTIONS

CAUTIONS

1. Avoid sharp knocks!

When carrying the microscope, or manipulating, handle the microscope gently taking care to avoid sharp knocks.

2. Locations of microscope

Avoid the following conditions: dust, vibration and exposure to high temperature, moisture or direct sunlight.

3. Dirt on the lens

Do not leave dust, dirt or finger marks on the lens surfaces. They will prevent you from clear observation of the specimen image.

MAINTENANCE

1. Cleaning the lenses

To clean the lens surfaces, remove dust using soft brush or gauze. Only for removing finger marks or grease, should soft cotton cloth, lens tissue or gauze lightly moistened with absolute alcohol (ethyl alcohol or methyl alcohol) be used.

Observe sufficient caution in handling alcohol.

2. Cleaning the plastic parts

Avoid the use of any organic solvent (for example, thinner, ether, alcohol, xylene etc.) for cleaning the plastic parts and painted surfaces of the instrument.

3. Never attempt to dismantle!

Never attempt to dismantle the instrument so as to avoid the possibility of imparing the operational efficiency and accuracy.

4. When not in use

When not in use, cover the instrument with the accessory vinyl cover, and store it in a place free from moisture and fungus.

5. Periodical checking

To maintain the performance of the instrument, we recommend to check the instrument periodically. (For details for this check, contact your dealer.)

NIKON CORPORATION

NOMENCLATURE AND USING

Evepieces

The eyepieces available as standard for the Microscope SMZ-2B are 10× and 15×. 20× and 33× are also available on order.

Note that no regular (biological) microscope eyepiece can be used interchangeably.

Diopter rings

Provided on both the righthand and lefthand eyepieces and adjustable to prevent out-of-focus during the time of zooming.

Microscope body

Microscope body clamp screw

For fastening firmly the main body of Microscope SMZ-2B onto the focus stand.

Episcopic illuminator

Incorporates a halogen lamp 6V— 10W, illumination angle is changeable. To achieve the most appropriate illumination for observation, adjust the position of the lamp.

Clips

To hold the specimen in place securely.

Stage plate

The black or opal surface is to be used according to the specimen to be observed. On the diascopic illuminator base, use the transparent class plate.

Zooming knob

Rotation of the zooming knob changes the magnifying power of the objective continuously from 0.8× to 5×, the powers being indicated on the knob.

Flexible arm attaching screws

Provided on both the right and left sides of the focus stand arm and used to attach the flexible arm.

Focus knobs

Rotation of the knobs moves the main body of microscope.

For increasing the tightness of rotation turn the right-and lefthand knobs in the opposite directions, i.e. in the fastening direction for each knob, as indicated by the arrow "TENSION" marked on the knobs.

The tightness will be decreased by releasing the each knob at the same time.

Transformer XN

MAGNIFICATION TABLE

Focus stand base

Focus stand arm

Objec- tive	Auxi- liary objec- tives	Work- ing dis- tance (mm)	Eyepieces							
			10× Field number 23		15X Field number 14		20× Field number 12		33× Field number 5	
				Real field (mm)	Total magnifi- cation	Real field (mm)	Total magnifi- cation	Real field (mm)	Total magnifi- cation	Real field (mm)
0.8~ 5×	_	100	8~50×	28.8~4.6	12~75X	17.5~2.8	16~100×	15~2.4	26.4~165X	6.3~1
	0.5×	181	4~25X	57.6~9.2	6~37.5×	35.0~5.6	8~50×	30~4.8	13.2~82.5X	12.6~2
	2X	32.7	16~100X	14.4~2.3	24~150X	8.75~1.4	32~200X	7.5~1.2	52.8~330×	3.1~0.5

OPERATING PROCEDURE

- 1. Place the specimen onto the stage plate.
- 2. Turn both diopter rings on the binocular tubes until the 0 line to the index.
- Looking into the eyepieces of the binocular tubes, adjust the interpupillary distance by moving the binocular tubes, so that both the circular viewfields are brought into coincidence.
- 4. Set the zooming knob to $5\times$.
- Rotating the focus knobs, bring the specimen image into focus. If, beforehand, the distance between the specimen and the bottom of the microscope main body is made about 100mm, focusing will be facilitated.
- 6. Rotate the zooming knob to $0.8\times$.
- Manipulating the right— and lefthand diopter rings one after the other, bring the images into focus independently.
- Repeat the above procedures from 4. to 7. to make sure of exact focusing, and, then the image will keep its sharpness all the time of continuous magnification changes by zooming.

We reserve the right to make such alterations in design as we may consider necessary in the light of experience.

For this reason, particulars and illustrations in this handbook may not conform in every detail to models in current production.

ACCESSORIES

Eyepieces:

 $20\times$ and $33\times$

Auxiliary objectives:

 $0.27\times$, $0.4\times$, $0.5\times$, $0.7\times$, 1.5× and 2.0×

(with $0.5\times$ objective, use Extension ring)

Universal illuminator

Fluorescent ring illuminator

Fiber optics illuminators:

Bifurcated and Ring illuminators

Diascopic illuminator base

Episcopic illuminator:

Incorporates a halogen lamp 6V-10W

(G.E. No. 787)

Flexible supporting arm

Universal arm

Universal stand:

Table stand and Table clamp stand

NIKON CORPORATION

9-16, Ohi 3-chome, Shinagawa-ku, Tokyo 140, Japan Tel: +81-3-3773-8122 Fax: +81-3-3773-8115

NIKON INC.

Instrument Group

1300 Walt Whitman Road, Melville, NY 11747-3064, U.S.A. Tel: +1-516-547-8500 Fax: +1-516-547-0306

NIKON EUROPE B.V.

Schipholweg 321, P.O. Box 222, 1170 AE Badhoevedorp, The Netherlands Tel: +31-20-44-96-222 Fax: +31-20-44-96-299

NIKON AG

Kaspar-Fennerstrasse 6, 8700 Küsnacht, Zürich, Schweiz Tel: +41-1-913-62-00 Fax: +41-1-910-37-44

NIKON GMBH

Tiefenbroicher Weg 26 40472 Düsseldorf 30, Fed. Rep. of Germany Tel: +49-0211-94-14-0 Fax: +49-0211-94-14-322

NIKON FRANCE S.A.

191, rue du Marché Rollay, 94504 Champigny Sur Marne Cedex France Tel: +33-1-45 16 45 16 Fax: +33-1-45 16 00 33

NIKON INSTRUMENTS S.P.A.

Via Tevere, 54 50019 Sesto Fiorentino (FI), Italia Tel: +39-55-3009601 Fax: +39-55-300993

NIKON U.K. LTD.

380 Richmond Road, Kingston, Surrey KT2 5PR, United Kingdom Tel: +44-181-541-4440 Fax: +44-181-541-4564

161: +44-181-541-444U Fax: +44-181-541-4564

ISO 9001 Certified

EVGI Conflicate No. 5257

NIKON CORPORATION Instruments Division Yokohama Plant

66 E 91.6.E.13

Printed in Japan